



Declaration of Conformity

Manufacturer:

ResMed Pty. Ltd.
1 Elizabeth Macarthur Drive
Bella Vista
NSW 2153
Australia

Authorised Representative:

ResMed SAS
Parc Technologique de Lyon
292 Allée Jacques Monod
69791 Saint Priest Cedex
France

Notified Body:

TÜV SÜD Product Service
GmbH
Ridlerstraße 65
80339 München
Germany

Product: AcuCare F1-1

Intended Use: The AcuCare F1-1 mask is a non-invasive accessory used for channelling airflow to a patient from a positive airway pressure (PAP) device.
The mask is:

- to be used on patients weighing >66 lb (30 kg), for whom non-invasive positive airway pressure therapy has been prescribed.
- a disposable device, intended for short-term treatment (up to 7 days) of a single patient, in the hospital environment only.
- intended to be used with breathing circuits or positive pressure ventilation (PPV) devices that provide

Classification: IIa according to Rule 2

GMDN: 57813 CPAP/BPAP face mask, single use

Conformity Assessment Route: Annex II (excluding Section 4), 93/42/EEC

We herewith declare that the above mentioned products are in conformity with the Council Directive 93/42/EEC for medical devices including the MDD amendment 2007/47/EC.

Compliance is applicable from the date listed below. All supporting documentation is retained at the premises of the manufacturer.

This declaration is issued under the sole responsibility of ResMed Pty. Ltd.

EC Certificate Number: G1 049861 0158

Signed at Sydney, Australia on: 18 June 2021

Johanna Wright
Director of Regulatory Affairs
ResMed Pty. Ltd.

EC159a

First issued: 3 March 2015